

PATIENT EDUCATION

Pole Mounted CADD Solis Administration Procedure (Vial Mate in Continuous Mode)

Properly administering your medication at home is important to your safety. In the event of an emergency, call 911.

SUPPLIES:

- Vial Mate bag of medication
- CADD tubing with in-line filter (change **DAILY** unless otherwise instructed)
- IV pole
- Cadd Solis Pump with Rechargeable battery
- CADD Solis pole clamp
- Charging cord
- 4-AA batteries for emergency backup only
- Alcohol/antiseptic wipes
- (2) Sodium Chloride (Saline) syringes
- (1) Heparin prefilled syringe (if indicated)

SASH FLUSH PROCEDURE:


- S** Saline ___ ml
- A** Administer medication as instructed by your nurse
- S** Saline ___ ml
- H** Heparin ___ ml (if required)

PROCEDURE:

1. Clean work area. Wash hands thoroughly for at least 20 seconds. Gather supplies.
2. Check medication label for name, drug, frequency, and expiration. Inspect the medication syringe for any cracks, leaks, particulate matter, and clarity of medication. Contact us for any discrepancies or concerns.
3. Prepare your prefilled syringes for flushing your IV catheter, as instructed by your nurse.
4. Prepare Vial Mate as follows:

PATIENT EDUCATION

Pole Mounted CADD Solis Administration Procedure (Vial Mate in Continuous Mode)



To mix, hold the white vial gripper firmly in one hand and the blue port adaptor in the other hand. Twist slightly to line up the ribs of the blue port adaptor and white vial gripper.



Push the blue port adaptor fully into the white gripper.




Hold bag over vial. Gently squeeze the solution until the vial is about half full.

PATIENT EDUCATION

Pole Mounted CADD Solis Administration Procedure (Vial Mate in Continuous Mode)



Fold vial up to rest against IV solution bag. While holding vial and bag together, shake to mix drug into solution.



Hold bag below the vial. Gently squeeze bag to force air bubble into the vial. Release IV solution bag to drain solution from the vial. Repeat until vial is empty. Make sure drug is completely mixed and the vial is empty.

5. Remove CADD tubing with in-line filter from package. Remove the blue clip from the top of the cassette by pulling up on the blue clip. Removing blue clip causes flow stop to close, preventing unintended gravity flow
6. Remove protective tab from entry port of medication bag.
7. Remove the cover from the spike on tubing and insert spike into medication bag using a pushing twisting motion. **DO NOT TOUCH SPIKE.**
8. Press the Power button on the right side of pump. Pump will make a series of beeps. Main screen will appear with pump mode showing and **"READY TO BEGIN"** in center of screen.

PATIENT EDUCATION

Pole Mounted CADD Solis Administration Procedure (Vial Mate in Continuous Mode)

9. Open cassette latch lever 90 degrees and attach new tubing to pump (hook hinged end first). Push up on the cassette until it firmly clicks into place OR place the pump upright on a firm, flat surface and then press down on the latch side of the pump until the cassette clicks into place. Close Latch.

***NOTE: Do not force latch. If you are unable to latch the cassette with minimal to no resistance, the cassette is not in the proper latching position. Unlatch the cassette and repeat the process. Top of screen will flash "High Volume or Standard Administration Set Latched".**

10. Check the pump screen, if **RESET RESERVIOR VOLUME TO _____? ML** appears, press **"YES"** on keypad. *****This question will not appear during initial infusion*****
11. Pump will then display **"PRIME TUBING?"** Press **"YES"** on keypad. Make sure any clamps on tubing are open and turn pump on its side (latch-side down) to assist with the removal of air bubbles. Pump screen will read **"Disconnect tubing from patient, open clamps then press Prime"**. Press **"PRIME"** on keypad, until a drop of fluid comes out.
12. If administering a partial dose, reset the reservoir volume at this time. Press **"TASKS"** on keypad. **"RESET RESERVIOR"** will be highlighted, press **"SELECT"** on keypad. Pump will read **"RESET RESERVOIR VOLUME TO ML?"** Press **'YES'**.
13. Attach CADD Solis pump to pole clamp on IV pole.
14. Clean the injection cap on the end of your IV catheter with a new alcohol/antiseptic wipe vigorously for 30 seconds and let air dry for 60 seconds.
15. Attach a prefilled Sodium Chloride/Saline syringe to the injection cap of your IV catheter and flush the IV catheter as instructed by your nurse.
16. Disconnect the syringe from the injection cap of your IV catheter and discard.
17. Remove protective cap from the end of the CADD tubing and attach tubing to the injection cap on the end of your IV catheter. ****Do NOT TOUCH END OF TUBING****

NOTE: After a short time, the display turns itself off to save power. To reactivate display screen, press any white button on the keypad.

18. When **"START PUMP?"** Displays on the pump screen, press **"YES"** on the keypad. This will start the infusion. The pump will display that it is running, and the green indicator light will flash.

PATIENT EDUCATION

Pole Mounted CADD Solis Administration Procedure (Vial Mate in Continuous Mode)

AT THE END OF THE INFUSION:

19. When infusion is completed, the pump will alarm and display “RESERVOIR VOLUME IS ZERO-PUMP STOPPED”. Press “ACKNOWLEDGE” to stop alarm.

NOTE: The tubing is to be changed DAILY. Reuse the same tubing when administering more than one infusion per day.

20. Turn pump off by pressing the Power button on the right side of pump.
21. Disconnect the tubing from the injection port of your IV catheter.
22. If another dose of medication is to be administered the **SAME** day, place a light blue protective cap on the end of the tubing, keeping it sterile.
23. Clean the injection cap on the end of your IV catheter with an alcohol/antiseptic wipe vigorously for 30 seconds and let air dry for 60 seconds.
24. Attach a prefilled Sodium Chloride/Saline syringe to the injection cap of your IV catheter and flush the IV catheter as instructed by your nurse.
25. Disconnect the syringe from the injection cap of your IV catheter and discard.

If your catheter requires the Use of Heparin: Complete steps 26 - 28, otherwise, skip to step 29.

26. Clean the injection cap on the end of your IV catheter with an alcohol/antiseptic wipe vigorously for 30 seconds and let air dry for 60 seconds.
27. Attach a prefilled Heparin syringe to the injection cap of your IV catheter and flush the IV catheter as instructed by your nurse.
28. Disconnect the syringe from the injection cap of your IV catheter and discard.
29. Flush any additional lumens as directed in your plan of treatment.
30. Remove medication bag/tubing or cassette from pump and discard if all infusions are completed for the day.
31. Wash hands thoroughly, for at least 20 seconds.

PATIENT EDUCATION

Pole Mounted CADD Solis Administration Procedure (Vial Mate in Continuous Mode)

**** All supplies and waste can be double bagged and disposed of with your household garbage****

NOTE: The CADD Solis pump runs on a rechargeable battery and will need plugged into a wall jack using the AC adaptor for 4-6 hours daily to recharge. In an emergency, the CADD Solis pump can run on 4-AA batteries. Please refer to the patient teaching handout for instructions of how to charge the CADD Solis pump or change the 4-AA batteries.

NOTE: If you are relying on battery power, and the yellow and green light on the top left corner of the pump starts blinking, and a “Battery Depleted” or “Battery Low” message appears on the display screen, you should change the batteries. Refer to CADD Solis Pump Disposable Battery Change Sheet.

NOTES:

CAREPATHrx

The information included in this document is intended for informational and educational purposes only. It is not intended to replace medical advice offered by a physician or other qualified health provider. If you think you are experiencing a medical emergency, call 911 immediately.

Copyright© 2021 by CarepathRx. All rights reserved.